

SESANS Reproducibility Working Group

Update and Activities


canSAS-XIV at SAS2024 (Taipei)—3 November 2024

Gregory Smith (ISIS)—Working Group chair

Wim Bouwman (TU Delft)—Presenter


Spin-echo small-angle neutron scattering

SESANS



Working Group members

- Wim Bouwman (TU Delft)
- Robert Dalgliesh (ISIS Neutron and Muon Source)
- Henrich Frielinghaus (Jülich Centre for Neutron Science, MLZ)
- Fankang Li (Oak Ridge National Laboratory)
- Andrew Parnell (The University of Sheffield)
- Steven Parnell (ISIS Neutron and Muon Source)
- Roger Pynn (Indiana University Bloomington)
- Gregory Smith (Chair, ISIS Neutron and Muon Source)



Group Activities

SESANS Reproducibility Working Group

[Page](#) [Discussion](#)

[View source](#) [History](#)

The spin-echo SANS (SESANS) working group is a subgroup of the [canSAS Reproducibility and Reliability Working Group](#).

Contents [hide](#)

- 1 Mission
- 2 Members
- 3 News and Status
- 4 SESANS Nomenclature
- 5 Data Repository
- 6 SESANS Instruments
- 7 Mailing List

- Wiki page with information about SESANS and links to instruments
 - https://wiki.cansas.org/index.php?title=SESANS_Reproducibility_Working_Group
 - Curation of data repository of possible SESANS standards
 - https://zenodo.org/communities/sesans_reproducibility_working_group/
 - Mailing list of SESANS-related updates
 - <https://www.jiscmail.ac.uk/cgi-bin/webadmin?SUBED1=SESANS-WORKING-GROUP&A=1>



Agreed nomenclature for SESANS data

Statement available on CanSAS wiki

- **SESANS data function of "spin-echo length" δ .**
No preferred unit, (μm , nm , or \AA) whatever suits natural length scale sample.
- **SESANS data measured "normalized scattering correlation function".**
Normalization measured polarization sample (P) with polarization reference (P_0).
 $\ln(P/P_0)/(t^* \lambda^2)$.
Instrument and wavelength independent data.
No recommended variable for the normalized scattering correlation function.
We recommend unit ($\text{cm}^{-1} \text{\AA}^{-2}$), natural units sample thickness (cm) and for neutron wavelength (\AA).


Cross-calibration of SESANS instruments

- Porous alumina membrane flexible graphite: ORNL, TU Delft, and ISIS (Larmor)



SESANS file format

- Text-based SES format with header and four-column data.
- Improve header with unequivocal identification of data and facility.
- What to include in future Nexus based file format.




```
88214_sesans_tcorr.ses
FileFormatVersion      1.0
DataFileName          {ORNL FlexiPor -88deg 1MHz Scruffy}_SESANS
Sample                88214
Thickness             0.060000
Thickness_unit         mm
Theta_zmax            0.1000000000000001
Theta_zmax_unit        radians
Theta_ymax            0.1000000000000001
Theta_ymax_unit        radians
Orientation           Z
SpinEchoLength_unit   A
Depolarisation_unit  A-2 cm-1
Wavelength_unit       A
Echo_constant         26.494816936180609

BEGIN_DATA
SpinEchoLength Depolarisation Depolarisation_error Wavelength
179.105 -0.49888 0.142202 2.6
207.719 -0.662651 0.118221 2.8
238.453 -1.01359 0.102436 3
271.307 -1.01216 0.092399 3.2
306.28 -1.12373 0.0850063 3.4
343.373 -1.41995 0.0790347 3.6
382.585 -1.58498 0.0738817 3.8
423.917 -1.79894 0.0683251 4
467.369 -2.07709 0.0623868 4.2
512.94 -2.39655 0.0591027 4.4
560.63 -2.64254 0.056531 4.6
610.441 -2.99566 0.0548285 4.8
649.189 -3.23433 0.076687 4.95
```

Data corrections

- Neutron wavelength and finite detector size impact experimentally determined data.
- Simple corrections using transmission.
- Desire: understand corrections. Reproducibility measurements different sources (mono vs ToF, different λ).

SCIENTIFIC REPORTS





OPEN

Data Correction of Intensity Modulated Small Angle Scattering

Fankang Li¹, Steven R. Parnell², Robert Dalgliesh³, Adam Washington³, Jeroen Plomp² & Roger Pynn^{1,4}

Standard samples for calibration

- Cross-calibration data for understanding SESANS measurements to get comparable data from different facilities.
- Currently, gratings with well-defined spacings (see below 0.5 μm).
- Need better options calibration standards (spin-echo length and depolarization). Ideally primary standards on length scales from $\sim 40 \text{ nm}$ to $\sim 40 \mu\text{m}$.



SESANS reproducibility working group

- Communication platform
 - Nomenclature and symbols
 - Cross calibration
 - Standard file format
-
- Data corrections
 - Standard samples